

Report For: Continental Filing System Inc.
2045, rue Berlier
Laval, Quebec
H7L 3M9
Phone: 450-686-2866 x202
Email: melkon@continental-fs.ca

Laboratory #: 855925-21
Report Date: February 18, 2021
Received Date: February 10, 2021

Attention: Melkon Iskenderian
Specimen: #1: Surgical Mask. Item#: 19022 MDEL: 13747 Manufactured by: CONTINENTALFS INC.
Lot#:220221. Expiration date: March 2023.


TEST REPORT

One specimen, consisting of face masks, was submitted to be tested for bacteria filtration efficiency, differential pressure, particle filtration efficiency, synthetic blood penetration and flame spread to determine barrier classification level as per ASTM F2100-20 requirements.



This report is subject to the following terms and conditions: 1. This report relates only to the specimen provided and there is no representation or warranty that it applies to similar substances or materials or the bulk of which the specimen is a part. 2. The content of this report is for the information of the customer identified above only and it shall not be reprinted, published or disclosed to any other party except in full. Prior written consent from Cambridge Materials Testing Limited is required. 3. The name Cambridge Materials Testing Limited shall not be used in connection with the specimen reported on or any substance or materials similar to that specimen without the prior written consent of Cambridge Materials Testing Limited. 4. Neither Cambridge Materials Testing Limited nor any of its employees shall be responsible or held liable for any claims, loss or damages arising in consequence of reliance on this report or any default, error or omission in its preparation or the tests conducted. 5. Specimens are retained 6 months, test reports and test data are retained 7 years from date of final test report and then disposed of, unless instructed otherwise in writing. 6. When making a statement of conformity to a specification or standard the report will make the statement of conformity based on the absolute value of the test result. Test Report Template Revision August 20, 2019

Per Stephen Brown Authorized By Stephen Brown
Per Diana Kalinowski Technician, Diana Kalinowski



Laboratory # 855925-21
Continental Filing System Inc

Medical Face Mask Packaging Requirements

Package Information	Packaging Displayed Information
Manufacturer Name	CONTINENTALFS INC
Product / Style Name	Surgical Mask/Masque Chirurgical
Lot Number	220221
Graphical representation indicating the performance level met with the technical requirements of the indicated performance level including a prominent visual indication of the performance level.	Yes
Requirements (Pass / Fail)	Pass

Note: ASTM F2100-20 requires verification of packaging, which prominently displays the above packaging information.

Medical Face Mask Material Requirements

Characteristic	Level 1 Barrier	Level 2 Barrier	Level 3 Barrier	Summary Results
Bacterial Filtration Efficiency, %	≥95	≥98	≥98	Pass any Level
Differential Pressure, mm H ₂ O/cm ²	<5.0	<6.0	<6.0	Pass any Level
Sub-Micron Particulate Filtration Efficiency at 0.1 micron, %	≥95	≥98	≥98	Pass any Level
Synthetic Blood Penetration minimum pressure in mmHg for pass result	80	120	160	Pass Level 2
Flame Spread	Class 1	Class 1	Class 1	Pass any Level
OVERALL PERFORMANCE LEVEL	Complete - Level 2			

Note: All five tests must be performed and meet with the requirements of ASTM F2100-20 in order to determine the final overall performance level of the mask, otherwise, the performance level is deemed, "Undetermined".



DIFFERENTIAL PRESSURE

EN 14683:2019 edition Annex C

Each specimen was conditioned for 4 hours minimum at 21+/-5 C and 85+/-5 % R.H.

RESULTS

<u>Specimen ID</u>	<u>Area ID</u>	<u>Differential Pressure (mmH2O/cm²)</u>
1-1	1	4.5
	2	4.9
	3	5.1
	4	5.2
	5	5.0
	AVERAGE	4.9
1-2	1	4.1
	2	4.3
	3	5.6
	4	4.6
	5	6.0
	AVERAGE	4.9
1-3	1	5.2
	2	5.2
	3	4.7
	4	5.1
	5	4.6
	AVERAGE	4.9
1-4	1	5.3
	2	4.8
	3	5.0
	4	4.7
	5	4.9
	AVERAGE	4.9
1-5	1	5.5
	2	5.0
	3	4.9
	4	4.7
	5	4.7
	AVERAGE	4.9

Mask Surface Area: 25mm diameter (x5 test areas) (4.9 cm²)

Air Flow Rate: 8 L/min

Mask Location Specimen taken from: 5 Areas from each specimen distributed all surface wide

Note: For a test plan of 5 specimens, no failure is allowed for an Acceptable Quality Limit of 4.0%.



SYNTHETIC BLOOD PENETRATION

ASTM F1862/F1862M-17 at 120 mmHg pressure


RESULTS

Specimen #	Test Pressure (mmHg)	Total Number of Specimens	Number of Pass Specimens
1	120	32	30

Note: Acceptable Quality Limit of 4.0% is met for single sampling plan when 29 or more of the 32 tested specimens show pass results.

Material construction type	Non-woven fabric
Supplier	CONTINENTALFS INC
Lot number	220221
Date of receipt	February 9, 2021
Date of test	February 17, 2021
Fluid velocity (cm/s)	555
Volume of impact fluid (ml)	2
Angle of pneumatic valve to horizontal	3°
Description target area mask	Outer blue ripple area (see Note)
Distance from tip cannula to mask (in)	12
Technique to enhance visual detection	Cotton swab used to lightly daub on the surface
Conditioning parameters	21±5°C, 85±5% R.H for minimum of 4 hours

NOTE: The outside surface of the mask is exposed to the blood stream in order to observe whether penetration occurred on the inner surface of the mask that could be contacting the wearer's face. Penetration on the inner facing of the mask constitutes failure (ASTM F1862/F1862M-17 section 4.2).



FLAME SPREAD


The specimen, consisting of 5 masks, was tested in accordance to 16 CFR 1610 (1-1-16 Edition).

	Specimen #	RESULT	CONCLUSION
Specimen #1	1-1	IBE	Classified as Class 1
	1-2	IBE	
	1-3	IBE	
	1-4	IBE	
	1-5	IBE	

IBE: Ignited but extinguished

- Test:** Flame Resistance 45° angle test. One-Second Flame Impingement.
- Type of fabric:** Without a raised fiber surface
- Surface tested:** Face
- Type of test:** Original State
- Direction tested:** Length
- Testing Conditioning:** Specimens conditioned at 105°C for 30 min, then placed in desiccator
- Requirements:** The flame spread time for textile products without a raised fibre surface must be greater than 3.5 seconds.

Note: For a test plan of 5 specimens, no failure is allowed for an Acceptable Quality Limit of 4.0%.



PARTICLE FILTRATION EFFICIENCY

Particles: Monodispersed polystyrene latex spheres (PSL)
Particles Counter: TSI scanning mobility particle sizer spectrometer 3082 and CPC
Tested as per ASTM F2299, non-neutralized aerosol challenge measured over 3 minutes (test specimen / control counts before and after test specimen and averaged)

Test Side: Inside
Area Tested: 21.7 cm²
Particle Size: 0.1 µm
Laboratory Conditions: 23°C, 38% relative humidity (RH)

Requirements ASTM F2100-19:

Particle filtration efficiency at 0.1 micron (%)
Level 1 Barrier: ≥95
Level 2 Barrier: ≥98
Level 3 Barrier: ≥98

RESULTS

Specimen #	Average Control Counts	Specimen Counts	Filtration Efficiency (%)	Specimen (Pass/Fail)	FINAL RESULT
1-1	226,465	658	<99	Pass	▲Pass any Level
1-2	221,250	1,194	99	Pass	
1-3	253,682	843	<99	Pass	
1-4	229,770	1,401	99	Pass	
1-5	192,025	745	<99	Pass	

▲Note: Tested under Laboratory#: 850549-20, reported December 17, 2020.

Note: The PFE equipment was outsourced and located at University of Toronto, 223 College Street, Toronto, ON, M5T 1R4.



BACTERIAL FILTRATION EFFICIENCY

A Bacterial Filtration Efficiency (BFE) test was completed according to the procedure in ASTM F2101-19 to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts recovered downstream. A suspension of *S. aureus* was aerosolized using a nebulizer and delivered to the test article at a constant rate with a target delivery rate of $1.7 \times 10^3 - 3.0 \times 10^3$ colony forming units (CFU) per test article with a mean particle size of $3.0 \pm 0.3 \mu\text{m}$. The aerosolized suspension was drawn through the test article which was clamped in a six stage Andersen air sampler, at a constant flow rate of 28.3 liters per minute (LPM), for collection on bacteriological agar plates.

Challenge Microbe: *Staphylococcus aureus* ATCC 6538
Test Side: User side facing challenge
Area Tested: $\sim 38.5 \text{ cm}^2$
Flow Rate: 28.3 LPM
Test Article Conditioning: $85 \pm 5\% \text{ RH}$ at $25.0 \pm 0.5^\circ\text{C}$ for a minimum of 4 hours
Challenge Level: $2.9 \times 10^3 \text{ CFU}$
Mean Particle Size: $2.9 \mu\text{m}$

Requirements ASTM F2100-19:
Bacterial filtration efficiency (%)
Level 1 Barrier: ≥ 95
Level 2 Barrier: ≥ 98
Level 3 Barrier: ≥ 98

RESULTS

Specimen #	Total CFU Recovered	Percent BFE (%)	Specimen (Pass/Fail)	FINAL RESULT
1-1	4	99.9	Pass	^Pass any Level
1-2	1	>99.9	Pass	
1-3	4	99.9	Pass	

^Note: Tested under Laboratory#: 850549-20, reported December 17, 2020.

The filtration efficiency percentages were calculated using the following equation:

$$\% \text{ BFE} = \frac{C - T}{C} \times 100$$

C = Challenge Level
T = Total CFU recovered downstream of test article

Note: Testing performed by GAP EnviroMicrobial Services Ltd., 1020 Hargrieve Road, Unit 14, London, Ontario, Canada, N6E 1P5